



## **Declaration of Conformity**

The following equipment is declared to be in conformity according to the following directives and international standards when the equipment is installed and used in a manner consistent with the manufacturer's recommendations and reference documents.

Manufacturer:

**Zenitel Norway AS** 

Post Box 4498 Nydalen

0403 Oslo Norway

**Product name:** 

SCAS3 HOSPITAL UNIT WALL W/2

**PUSHBUTTONS** 

**Product number:** 

3006102064

**Electromagnetic Emissions:** 

EN61000-6-3(2001) + A11(2004)

(by Council Directive 2004/108/EC)

(by Council Directive 2004/108/EC)

**Electromagnetic Immunity:** 

EN61000-6-1(2001)

The technical file is maintained at the company address listed above.

The manufacturer is solely responsible for making this declaration.

ZENITEL NORWAY AS

Sandakerveien 24 C Postboks 4498 Nydalon

TEL + 47 21 51 25 0

Name, Surname

**Position/Title** 

Place of issue

Date of issue

Bjørn Grading

**Quality & Logistics Director** 

Horten

2013-07-19